



Occupational radiation monitoring practices in the southwest region of Pakistan

Sunnia Shafiq^{1}, Hafiz Khush Naseeb Ahmad², Hameeda Naheed³, Hina Manzoor⁴, Jamila Shuja³, Awais-ur Rehman¹, Imran Arif², Muhammad Sadiq², Shagufta Kanwal², Shehla Iftikhar⁵*

¹ MS CENAR Hospital Quetta, Pakistan

² MBBS, MS CENAR Hospital Quetta, Pakistan

³ MBBS, DMRD CENAR Hospital Quetta, Pakistan

⁴ MSc CENAR Hospital Quetta, Pakistan

⁵ MBBS, MCPS, FCPS, EDIR CENAR Hospital Quetta, Pakistan

*Corresponding author: Sunnia Shafiq, Address: CENAR Staff Colony, near BMC, Brewery Road Quetta, Pakistan

Email: Sunniyakhan19@gmail.com, Tel: +923336362626

Abstract

Background & Aims: The article aims to evaluate trends in occupational radiation exposure for employees in different departments at the Center for Nuclear Medicine and Radiotherapy (CENAR), Quetta, Pakistan.

Materials & Methods: This retrospective study examination of 9780 film dosimeter dose data was carried out from 2019 to 2023. The quantities analyzed were the highest and minimum annual effective doses, distribution of workers, and their average yearly effective doses in different effective dose intervals. The annual collective effective dose and average effective dosage were also included.

Results: Yearly average effective doses in Radiotherapy, Nuclear Medicine, Medical Physics, and Diagnostic Radiology range from 0.5 to 3.00 mSv. Records show that 90% of workers fall within 1-3.00 mSv, with 10% between 0.5 mSv and the minimum detectable level. Minimum and maximum yearly individual doses in Radiotherapy and Nuclear Medicine are 0.10 mSv and 2.3 mSv. Average doses for the fields range from 1.01 to 1.14 mSv.

Conclusion: Employees received doses that were lower than the yearly maximum, showing successful radiation shielding. The findings imply that safety measures are followed in the researched departments, which is important for clinical practice. This research improves our understanding of occupational radiation exposure, notably at CENAR, and promotes radiation safety in medical settings.

Keywords: Annual Dose, Distance Shielding, Occupational Radiation, Radiation Protection, Radioactive Sources, Time

Received 03 June 2024; accepted for publication 07 July 2024

Introduction

The use of radioactive sources spans various fields, including agriculture, industry, research, medicine, and

education, providing numerous benefits. Advances in healthcare and an aging population have increased the use of radiation and radionuclides for diagnostic and

therapeutic purposes (1). While the general public is typically exposed to low-level background radiation from natural sources, medical equipment such as nuclear medicines and X-rays—both industrial and diagnostic—constitute the primary sources of man-made radiation exposure, accounting for approximately 4% and 11% of exposure respectively (2).

Most exposure to low radiation doses occurs through medical and occupational contexts involving ionizing radiation. Research indicates that cancer risk correlates directly with the absorbed radiation dose, underscoring the importance of developing models to understand the carcinogenic effects of low-level exposure (3). The International Commission on Radiological Protection (ICRP) aims to provide a comprehensive system and practical guidelines for radiation protection, addressing exposure controls across medical, occupational, and environmental settings to prevent radiological incidents without hindering beneficial practices (4).

Occupational exposure refers to the exposure of workers to ionizing radiation in their workplace, excluding exposures not covered by standards or those exempt from regulation (5). For workers exposed to medical radiation sources, adherence to the guidelines in the International Basic Safety Standards for Protection against Ionizing Radiation and the Safety of Radiation Sources is crucial. Government agencies and organizations must consider dose calculations for radiation workers when assessing radiation hazards and implementing protective measures (6).

The principle of Time-Distance-Shielding (TDS) is central to radiation safety, aiming to protect workers and the public from man-made radiation sources. If reducing exposure time is not feasible, maintaining a safe distance from the source can minimize exposure. When neither time reduction nor distancing is possible, proper shielding between the source and the subject is essential. Various devices and materials, such as lead glass, concrete walls, and lead bricks, are used to shield against unnecessary radiation (7).

In the southwest region of Pakistan, Quetta—the capital of Baluchistan—hosts the Centre for Nuclear Medicine and Radiotherapy (CENAR), the sole cancer

management center in the province. CENAR is dedicated to providing exceptional cancer care and ensuring patient satisfaction. It employs both sealed and unsealed radiation sources, as well as man-made radiation-generating machines, for therapeutic and diagnostic purposes. In the nuclear medicine department, Tc 99m and I-131 are used for routine diagnostics and treating thyroid conditions, while X-ray and mammography units are used in the radiology department. The radiotherapy department utilizes Linear Accelerators (LINAC) and Co-60 for cancer treatments.

Monitoring and measuring radiation is crucial for ensuring appropriate protection. One of the primary tools for personal dosimetry is the film badge, which monitors employees in radiation-prone areas. The film badge, containing a photographic film within a plastic holder, differentiates between beta, X-ray, gamma, and neutron radiations, and assesses individual doses through open windows and various filters (8). Developing the film from these badges reveals the optical density, total dose, and contributions from each type of radiation (9).

Adhering to standards set by the International Atomic Energy Agency (IAEA) and the International Commission on Radiological Protection (ICRP), the Pakistan Nuclear Regulatory Authority (PNRA) has established a permissible occupational dose level of 20 mSv for radiation workers in Pakistan (10). This study compares the radiation doses received by CENAR employees to the limits established by national and international regulatory agencies.

Materials & Methods

This observational research was conducted at CENAR in Quetta to monitor occupational radiation doses. Each employee handling radioactive material at CENAR is given a film badge with a unique identifying number. Each month, PINSTECH (Pakistan Institute of Nuclear Science and Technology) in Islamabad delivers the film badges. The measuring of an employee's occupational radiation exposure is advised for radiation workers using a film badge as part of their personnel monitoring system (9).

In this study, the radiation dose received in the year 2019 (January-December 2019) and the total dose expected in the last five years (January 2019- December 2023) by 35 radiation workers of CENAR, Quetta were assessed. To assess the workers' history of radiation exposure, the readings from the film badges that PINSTECH gave have been retained as records. Data analysis involved department-specific evaluation of radiation doses, comparison with established safety limits, and assessment of trends over the five years. The study additionally examined compliance with annual radiation exposure limits. While the descriptive nature of the study design provides valuable insights, it is

limited in establishing causal relationships and generalizability of findings beyond the observed data. The relatively small sample size of 35 participants necessitates further investigation with a larger cohort to enhance the generalizability of the results.

Results

For staff exposure limits to remain below the levels recommended by the International Commission on Radiological Protection (ICRP) in 1991 (as amended), personnel radiation monitoring is necessary (11). The total number of radiation workers working in CENAR, Quetta mentioned in Figure-1.




Fig. 1. Total Number of radiation workers from 2019 to 2023

Department wise Distribution among the years 2019 to 2023 is shown in Figure-2.




Fig. 2. Department Wise Distribution from 2019 to 2023

Departmental wise average dose distribution is shown as in Table-1.

Table 1. Departmental wise average dose distribution

Department	Year	Total Workers	Total Dose	Average Dose of Worker (mSv)	percent of Permissible Limit i.e., (20 mSv)
Radiotherapy	2019	14	16.8	1.2	6
	2020	15	15.45	1.03	5.15
	2021	14	13.02	0.93	4.65
	2022	14	14.14	1.01	5.05
	2023	13	15.6	1.2	6
Nuclear Medicine	2019	5	6.2	1.24	6.2
	2020	7	6.93	0.99	4.95
	2021	8	9	1.125	5.625
	2022	10	10	1	5
	2023	8	9.6	1.2	6
Radiology	2019	2	2.4	1.2	6
	2020	2	2.4	1.2	6
	2021	3	2.7	0.9	4.5
	2022	4	4.3	1.075	5.375
	2023	4	4.8	1.2	6
Medical Physics	2019	3	3.6	1.2	6
	2020	4	4.12	1.03	5.15
	2021	4	4.6	1.15	5.75
	2022	5	5.8	1.16	5.8
	2023	5	6	1.2	6
Others	2019	5	6	1.2	6
	2020	5	6	1.2	6
	2021	4	3.6	0.9	4.5
	2022	4	4.64	1.16	5.8
	2023	4	4.8	1.2	6

Discussion

The majority of radiation workers of CENAR received the annual doses less than 6% of the annual dose as shown in Table-1. The average occupational data in medical physics, radiotherapy, nuclear medicine and radiology in CENAR, Quetta between 2019 and 2023 were reported as shown as a figure in the Figure-1 and Figure-2. According to the Figure-1 and Figure-2, the data ranged from 1.03 mSv to 1.2 mSv, 0.93 mSv to 1.2 mSv, 0.99 mSv to 1.2 mSv and 0.90 mSv to 1.2 mSv respectively.

The average occupational exposure data of workers in nuclear medicine, radiotherapy, and diagnostic radiology in Pakistan between 2003 and 2007 was reported by Jabeen A et al. ranged from 1.39 mSv to 1.80 mSv, 1.05 mSv to 1.45 mSv, and 1.22 mSv to 1.71 mSv, respectively (12).

The annual doses of radiation workers in the nuclear medicine, radiotherapy, and diagnostic radiology divisions were reported by Weizhang W et al. These doses fell within the ranges of 1.2-1.6, 1.0-1.5, and 1.5-2.2 mSv, respectively (13).

According to research by Korir GK et al., 81% of radiation workers received doses in a year ranging from 1 mSv to 5 mSv, while 17% of workers had yearly doses of less than 1 mSv (14).

According to Careiro JV and Avelar R, 97.8% of workers' annual occupational exposures fell within the range of 5 mSv (15).

The results of our study are significantly lower than the suggested international yearly limit for one year (20 mSv) as recommended by ICRP (11) and for a period of 5 consecutive years which were also indicated by Jabeen A et al. (12).

When comparing the results of this study to previous studies, it is clear that there are variations in the radiation doses received by workers in different countries and at different times. For instance, the average occupational exposure data from Jabeen et al. (2003-2007) for medical physics, radiotherapy, and nuclear medicine in Pakistan ranged from 1.05 to 1.80 mSv (12), which is somewhat higher than the ranges observed in our study at CENAR Quetta (0.90 to 1.20 mSv) from 2019 to 2023. This difference could be due to improved radiation protection measures and stricter enforcement of safety protocols over the years. Furthermore, Weizhang et al. reported annual doses in China from 1.2 to 2.2 mSv for similar categories, which are higher than our findings (13). Differences in the types of equipment used, operational procedures, and adherence to safety standards might account for these variations.

Conclusion

This study's primary goal was to assess radiation doses received by radiation personnel in accordance with allowable limits (10, 11). Based on the annual and last five years' dose data of all CENAR Quetta radiation workers, it can be concluded that all radiation workers' radiation doses fell within the acceptable range set by national (10) and international (11) organizations. This confirms the sufficiency of the facilities for radiation protection and demonstrates the rationality of radiation protection techniques (16). The results demonstrate the effectiveness of radiation protection measures and the adherence to safety recommendations in the

departments under study, making them clinically meaningful and important. For the purpose of protecting the health of healthcare professionals who are frequently exposed to ionizing radiation at work, this knowledge is essential. The study's findings might also help with ongoing initiatives to optimize radiation exposure in medical settings and guide future modifications to radiation safety procedures. This research contributes to our comprehension of radiation exposure at work in a particular healthcare environment by providing information on radiation levels, trends, and the efficiency of radiation shielding techniques. The data offered is helpful to the larger area of radiation safety in medical practice in addition to being pertinent to the specific context of CENAR.

Limitations and Future Research

Future research should aim to include a larger sample size to improve the representativeness of the findings. Further studies could also explore the long-term health outcomes of radiation exposure among these workers to provide a more comprehensive understanding of the risks involved.

Acknowledgments

The author is cordially thankful to Dr. Hudair Samad for helping me throughout this research.

Ethical statement

The Ethical Review Committee has reviewed and approved the ethical statement associated with the manuscript titled "Occupational Radiation Monitoring Practices in the Southwest Region of Pakistan." This study complies with ethical guidelines, including obtaining informed consent, ensuring confidentiality, and managing risks, maintaining data integrity, and disclosing conflicts of interest. This approval, letter number Dir-3(17)/24, confirms adherence to the highest ethical standards and supports the publication of your paper.

Data availability

Data will be available as per request.

Author contributions

All Authors contributed equally.

Conflict of interest

The Authors declare no conflict of interest.

Funding/support

Nil

References

- Singh B, Singh J, Kaur A. Applications of radioisotopes in agriculture. *Int J Biotechnol Bioeng* 2013;4:167-74.
- Laidler JB. National council on radiation protection and measurements. NCRP report no. 70 nuclear medicine-factors influencing the choice of radionuclides in diagnosis and therapy. Bethesda, Maryland: NCRP Publications; 1982. 180 p.
<https://doi.org/10.1002/jlcr.2580200715>
- Charles M. Effects of Ionizing Radiation: United Nations Scientific Committee on the Effects of Atomic Radiation: UNSCEAR 2006 Report, Volume 1-Report to the General Assembly, with Scientific Annexes A and B.
<https://doi.org/10.18356/7fb405cb-en>
- Leszczynski D. The Grand Challenge: use of a new approach in developing policies in the area of radiation and health. *Front Pub Health* 2014;2:50.
<https://doi.org/10.3389/fpubh.2014.00050>
- Weizhang W, Wenyi Z, Ronglin C, Liang'an Z. Occupational exposures of Chinese medical radiation workers in 1986-2000. *Rad Protect Dosimet* 2005;117(4):440-3. <https://doi.org/10.1093/rpd/nci312>
- Mora P, Acuña M. Assessment of medical occupational radiation doses in Costa Rica. *Rad Protect Dosimet* 2011;147(1-2):230-2. <https://doi.org/10.1093/rpd/ncr341>
- Bolus NE. Review of common occupational hazards and safety concerns for nuclear medicine technologists. *J Nucl Med Technol* 2008;36(1):11-7.
<https://doi.org/10.2967/jnmt.107.043869>
- Martin A, Harbison SA. The external Radiation hazard. In: Martin A, Harbison SA, editors. *Introd Radia Protec* 1980; 85-106. https://doi.org/10.1007/978-1-4899-6940-8_8
- Bushong SC. A Book Report. *Radiol Technol* 2013;84(4):405-9.
- Authority PN. Pakistan Nuclear Regulatory Authority. International Commission on Radiological Protection. ICRP Publication 60: *Annals of ICRP* 21(1-3), New York. International Commission on Radiological Protection; 1991. Available from:
<https://www.icrp.org/publication.asp?id=ICRP%20Publication%2060>
- Jabeen A, Munir M, Khalil A, Masood M, Akhter P. Occupational exposure from external radiation used in medical practices in Pakistan by film badge dosimetry. *Rad Protect Dosimet* 2010;140(4):396-401.
<https://doi.org/10.1093/rpd/ncq134>
- Weizhang W, Wenyi Z, Ronglin C, Liang'an Z. Occupational exposures of Chinese medical radiation workers in 1986-2000. *Rad Protect Dosimet* 2005;117(4):440-3. <https://doi.org/10.1093/rpd/nci312>
- Korir GK, Wambani JS, Korir IK. Estimation of annual occupational effective doses from external ionising radiation at medical institutions in Kenya. *SA J Radiol* 2011;15(4). <https://doi.org/10.4102/sajr.v15i4.353>
- Carreiro JV, Avelar R. Occupational exposure in medical and paramedical professions in Portugal. *Rad Protect Dosimet*. 1991;36(2-4):233-6.
<https://doi.org/10.1093/oxfordjournals.rpd.a081003>
- Ohagwu AO, Njoku J. Evaluation of personnel radiation monitoring in radiodiagnostic centres in South Eastern Nigeria. *Afr J Basic Appl Sci* 2010;2(1-2):49-53.

This is an open-access article distributed under the terms of the [Creative Commons Attribution-noncommercial 4.0 International License](https://creativecommons.org/licenses/by-nc/4.0/) which permits copy and redistribute the material just in noncommercial usages, as long as the original work is properly cited.